


RESTRICTED

TECHNICAL SPECIFICATIONS OF THERMOMETER RESISTANCE

Ser	Particulars	Description		
1.	Name of the Equipment	Thermometer Resistance		
2.	Quantity	02 (Two)		
3.	Purpose	Monitoring high-precision air temperature & record maximum & minimum air temperature.		
4.	Type	To be mentioned		
5.	Model	To be mentioned		
6.	Standard	To be Mentioned		
7.	Technical Parameters	a.	Measuring Range	0~50°C
		b.	Resolution	Centigrade degree: 0.1 ⁰
		c.	Accuracy	:±(0.05%+1digits)
		d.	Display	Digital LCD/OLED display
		e.	Power Supply	AC&DC or 1.5v AA battery
		f.	Output	Read & Record air temperature values.
		g.	Sampling time	0.5 second
8.	Country of Manufacture	USA, CANADA, UK, EU COUNTRIES, JAPAN		
9.	Year of Manufacture	2022 or Later		
10.	Port of Shipment	To be mentioned		
11.	Manufacturer's Name and Address	To be mentioned		
12.	Supplier/ Local Agent	To be mentioned		
13.	Certificate and Document of Authentication	a. OEM Certificate b. Authorized Distributor Certificate. c. Authenticity and Genuineness certificate d. Warranty Certificate		
14.	Packing	Need to be packed		
15.	Warranty	At least 12 months from date of acceptance		
16.	Installation	Installation work has to be done by supplier		
17.	Test, Trial and acceptance	Acceptance will be carried out by CINS		
18.	Training	Training as required		
19.	Accessories	Display & Data processing Unit including cable and any other accessories required for full operation.		
20.	Other	Including all other standard accessories required for operation of the system.		
21.	Picture			

RESTRICTED

1 *[Signature]*
09/04/2024